

**AMENDMENT NO. 6 AUGUST 2021**

**TO**

**IS 7454 : 1991 REHABILITATION EQUIPMENT —  
WHEELCHAIRS, FOLDING, ADULT SIZE —  
SPECIFICATION**

*( First Revision )*

*(Page 3, clause 6.8.1)* — Substitute the following for the existing clause:

**‘6.8.1** The hand rims shall be fabricated from steel/aluminium tubing with minimum outside diameter of 17 mm.

NOTE — Hand rims are not required for Type 1 wheelchair.’